



AURELIA® TRANSFORM

3.2mil, Nitrile Powder Free Examination Gloves

AURELIA® GLOVES
A SUPERMAX HEALTHCARE BRAND

Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1184: 2002. Materials and articles in contact with foodstuffs.



10 x 200




XS	S	M	L	XL
98895	98896	98897	98898	98899

Sizes & Dimensions

Single wall thickness

Physical Properties



XS 70+10 mm
S 80+10 mm
M 95+10 mm
L 110+10 mm
XL > 110 mm

min. 240 mm
min. 240 mm
min. 240 mm
min. 240 mm
min. 240 mm

① min. 0.10 mm
② min. 0.06 mm
③ min. 0.05 mm

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 500%	Min. 400%
Strength at Break	Min. 6N	Min. 6N



Nitrile



Powder Free



5 Years



Finger Textured



3.6g +/- 0.2g



Ice Blue

