dexipro®

Blue Nitrile | 3.5g | Finger Textured | Powder Free | AQL 1.5

EN ISO 374-5:2016	LEVEL
Protection against Bacteria and Fungi	PASS
Protection against Viruses	PASS

100 gloves per box

10 boxes per case

120 cases per pallet

Medical Device Regulation (UK) 2002 (as amended)	CLASS I
Medical Device Regulation (EU) 2017/745	CLASS I
PPE Regulation (EU) 2016/425	CATEGORY III

EN ISO 374-1:2016+A1:2018	LEVEL	EN ISO 374-4:2019 Degradation %
40% Sodium Hydroxide (K)	6	1.4
30% Hydrogen Peroxide (P)	6	24.0
37% Formaldehyde (T)	6	10.9

Break Force	
Unaged	Aged
Min. 6N	Min. 6N

Tensile Strength	
Unaged	Aged
Min. 15.0 MPa	Min. 14.0 MPa

Elongation	
Unaged	Aged
Min. 500%	Min. 400%

EN ISO 374-1:2016+A1:2018 Chemical Permeation	1	2	3	4	5	6
Breakthrough Time (Minutes)	>10	>30	>60	>120	>240	>480

SMALL	MEDIUM	LARGE	X-LARGE
95896	95897	95898	95899

Single Wall Thickness (mm)	Width (mm)	Length (mm)
0.10±0.02	S 80-90	min. 240
0.07±0.02	M 90-10	
min. 0.05	L 100-110	
	XL 110-120	

Regulations	PPE Regulation (EU) 2016/425
	Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems
	ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms
	EN 21420: 2020 Protective gloves. General requirements and test methods.
	BS EN 1186: 2002. Materials and articles in contact with foodstuffs.
	BS EN 455 1-4 Medical gloves for single use.

dexipro.co.uk

MD | CE 2777 | LATEX | NON STERILE | EN ISO 374-5:2016 Virus | EN ISO 374-1:2016 + A1 2018/Type B KPT | AQL 1.5 | 21 PAP