




Product Specification – Motokit Single Use Breathalyser Box of 2

Product Information

REF:	3076
Description:	Motokit Single Use Breathalyser Box of 2
Manufacturer:	Test&Drive
Address:	Dekoracyjna 3/B38, 65-155 Zielona Gora, Poland
Characteristics:	<ul style="list-style-type: none">- Easy to use.- Tests exhaled breath.- Accurate results in just 2 minutes.- 2 tests in pack.- Colour of crystals changes based on level of alcohol concentration.
Sensitivity:	At an alcohol level of 0.2 promille the device showed a test sensitivity of 96%, with an alcohol level of 0.5 promille the device has a sensitivity of 98.6%, at 0.8 promille the device demonstrates test sensitivity of 100%.
Safety:	Test & Drive does not contain potassium dichromate, a poisonous and carcinogenic compound used in older generation breathalysers.



The information contained in this product specification is accurate for the above revision date.
However, components are subject to change without notice, when another specification will be issued.

Document Title:
Product Specification:
Product References:
Revision:

Product Specification
Motokit Single Use Breathalyser Box of 2
3076
V4 (05.26)



Additional Information

Technology:	White ReactionsRedox
Measuring Method:	Colour change
Measuring Colour:	White to red
Reading Time:	2 min
Accuracy:	0.1 promille
Measuring method:	Exhaled air
Operating temperature:	-25 to 40 degrees Celsius
Measurement ranges:	0.0, 0.2, 0.5, 0.8 (per mille)
Device dimensions:	60mm x 11mm

Product Dimensions

UNIT	Dimensions (LxWxH)	Weight
Each	140mm x 75mm x 20mm	19g
Pack of 30	180mm x 120mm x 120mm	600g

*The information contained in this product specification is accurate for the above revision date.
However, components are subject to change without notice, when another specification will be issued.*



Document Approval

RM-020162-PS - Ver 1

PRODUCT SPEC - 3076 - 5906874789056 - Motokit Single Use Breathalyser Box of 2
Author: Toni-Mai O'Brien

Approval Detail:

Approver	Location/Time
Toni-Mai O'Brien	Talke, UK - Tue 19th May 2026 - 14:52
Role	ID/Checksum
QA	12ab64ce-85ad-47ed-873e-5434980f9f1b PscyQPNA55ufejwSAU8fipuIiZU
Meaning	
Approved	

Approval History:

Version	Approval Date	Approved By	Checksum
1	Tue May 19 2026	Toni-Mai O'Brien	b54d27ed3a6c65af4db1a3bdb56bf3a

